

EDGE® Radiosurgery System

Bibliography

Snyder KC, Cunningham J, Huang Y, Zhao B, Dolan J, Wen N, Chetty IJ, Shah MM, Siddiqui SM. Dosimetric Evaluation of Fractionated Stereotactic Radiation Therapy for Skull Base Meningiomas Using HyperArc and Multicriteria Optimization. *Adv Radiat Oncol*. 2021 Feb 6;6(4):100663. Henry Ford Health System, Detroit, MI

Dai ZT, Ma L, Cao TT, Zhu L, Zhao M, Li N. Dosimetric and radiobiological comparison of treatment plan between CyberKnife and EDGE in stereotactic body radiotherapy for pancreatic cancer. *Sci Rep*. 2021 Feb 18;11(1):4065. Chinese Academy of Medical Sciences and Peking Union Medical College, Beijing, China

Bossart E, Mellon EA, Monterroso I, Elsayyad N, Diwanji T, Samuels S, Dogan N. Assessment of single isocenter linear accelerator radiosurgery for metastases and base of skull lesions. *Phys Med*. 2021 Jan;81:1-8 2021 Jan;81:1-8 Sylvester Comprehensive Cancer Center, University of Miami, Miami, FL

Chen L, Bai S, Li G, Li Z, Xiao Q, Bai L, Li C, Xian L, Hu Z, Dai G, Wang G. Accuracy of real-time respiratory motion tracking and time delay of gating radiotherapy based on optical surface imaging technique. *Radiat Oncol*. 2020 Jul 10;15(1):170. West China Hospital, Sichuan University, Chengdu, China

Akino Y, Mizuno H, Isono M, Tanaka Y, Masai N, Yamamoto T. Small-field dosimetry of TrueBeam™ flattened and flattening filter-free beams: A multi-institutional analysis. *J Appl Clin Med Phys*. 2020 Jan;21(1):78-87 Osaka University Hospital, Suita, Osaka, Japan

Paganini L, Reggiori G, Stravato A, Palumbo V, Mancosu P, Lobefalo F, Gaudino A, Fogliata A, Scorsetti M, Tomatis S. MLC parameters from static fields to VMAT plans: an evaluation in a RT-dedicated MC environment (PRIMO). *Radiat Oncol*. 2019 Dec 2;14(1):216. Humanitas Clinical and Research Center, Rozzano, (Milan), Italy

Galavis PE, Hu L, Holmes S, Das IJ. Characterization of the plastic scintillation detector Exradin W2 for small field dosimetry. *Med Phys*. 2019 May;46(5):2468-2476 New York University, Langone Medical Center, New York, NY

Covington EL, Snyder JD, Wu X, Cardan RA, Popple RA. Assessing the feasibility of single target radiosurgery quality assurance with portal dosimetry. *J Appl Clin Med Phys*. 2019 May;20(5):135-140 University of Alabama - Birmingham, South Birmingham, AL

Parsons D, Church C, Syme A. Toward a pre-clinical irradiator using clinical infrastructure. *Phys Med*. 2019 Feb;58:21-31 University of Texas Southwestern Medical Center, Dallas, TX

Yao S, Zhang Y, Chen T, Zhao G, Hu Z, Lu X, Liu Y. Dosimetric Comparison Between Jaw Tracking and No Jaw Tracking in Intensity-Modulated Radiation Therapy. *Technol Cancer Res Treat*. 2019 Jan 1;18:1533033819841061. Shanghai General Hospital, Shanghai, China

Mao W, Gardner SJ, Snyder KC, Wen NW, Zhong H, Li H, Jackson P, Shah M, Chetty IJ. On the improvement of CBCT image quality for soft tissue-based SRS localization. *J Appl Clin Med Phys*. 2018 Nov;19(6):177-184. Henry Ford Health System, Detroit, MI

Woon WA, Ravindran PB, Ekayanake P, Lim YYF. Validation of Delivery Consistency for Intensity-Modulated Radiation Therapy and Volumetric-Modulated Arc Therapy Plans. *J Med Phys*. 2018 Apr-Jun;43(2):119-128. The Brunei Cancer Center, Brunei Darussalam

Reggiori G, Stravato A, Mancosu P, Lobefalo F, Paganini L, Zucconi F, Palumbo V, Gaudino A, Scorsetti M, Tomatis S. Small field characterization of a Nanochamber prototype under flattening filter free photon beams. *Phys Med*. 2018 May;49:139-146. Humanitas Clinical and Research Hospital, Milan-Rozzano, Italy

Hillman Y, Kim J, Chetty I, Wen N. Refinement of MLC modeling improves commercial QA dosimetry system for SRS and SBRT patient-specific QA. *Med Phys*. 2018 Apr;45(4):1351-1359 Henry Ford Health System, Detroit, MI

Qin Y, Gardner SJ, Kim J, Huang Y, Wen N, Doemer A, Chetty IJ. Technical Note: Evaluation of plastic scintillator detector for small field stereotactic patient-specific quality assurance. *Med Phys*. 2017 Oct;44(10):5509-5516. Henry Ford Health System, Detroit, MI

Reggiori G, Stravato A, Pimpinella M, Lobefalo F, De Coste V, Fogliata A, Mancosu P, De Rose F, Palumbo V, Scorsetti M, Tomatis S. Use of PTW-microDiamond for relative dosimetry of unflattened photon beams. *Phys Med.* 2017 Jun;38:45-53. Humanitas Clinical and Research Hospital, Milan-Rozzano, Italy

Oliver JA, Kelly P, Meeks SL, Willoughby TR, Shah AP. Orthogonal image pairs coupled with OSMS for noncoplanar beam angle, intracranial, single-isocenter, SRS treatments with multiple targets on the Varian Edge radiosurgery system. *Adv Radiat Oncol.* 2017 Apr 26;2(3):494-502. UF Health Cancer Center -Orlando Health, Orlando, FL

Oliver JA, Zeidan OA, Meeks SL, Shah AP, Pukala J, Kelly P, Ramakrishna NR, Willoughby TR. The Mobius AIRO mobile CT for image-guided proton therapy: Characterization & commissioning. *J Appl Clin Med Phys.* 2017 May;18(3):130-136. UF Health Cancer Center -Orlando Health, Orlando, FL

Xu H, Brown S, Chetty IJ, Wen N. A Systematic Analysis of Errors in Target Localization and Treatment Delivery for Stereotactic Radiosurgery Using 2D/3D Image Registration. *Technol Cancer Res Treat.* 2017 Jun;16(3):321-331. Henry Ford Health System, Detroit, MI

Ravindran PB. A study of Winston-Lutz test on two different electronic portal imaging devices and with low energy imaging. *Australas Phys Eng Sci Med.* 2016 Sep;39(3):677-85. The Brunei Cancer Center, Brunei Darussalam

Mancosu P, Fogliata A, Stravato A, Tomatis S, Cozzi L, Scorsetti M. Accuracy evaluation of the optical surface monitoring system on EDGE linear accelerator in a phantom study. *Med Dosim.* 2016 Summer;41(2):173-9. Humanitas Clinical and Research Hospital, Milan-Rozzano, Italy

Tanny S, Sperling N, Parsai EI. Correction factor measurements for multiple detectors used in small field dosimetry on the Varian Edge radiosurgery system. *Med Phys.* 2015 Sep;42(9):5370-6. University of Toledo Medical Center, Toledo, OH

Wen N, Li H, Song K, Chin-Snyder K, Qin Y, Kim J, Bellon M, Gulam M, Gardner S, Doemer A, Devpura S, Gordon J, Chetty I, Siddiqui F, Ajlouni M, Pompa R, Hammoud Z, Simoff M, Kalkanis S, Movsas B, Siddiqui MS. Characteristics of a novel treatment system for linear accelerator-based stereotactic radiosurgery. *J Appl Clin Med Phys.* 2015 Jul 8;16(4):125-148. Henry Ford Health System, Detroit, MI

Kielar KN, Mok E, Hsu A, Wang L, Luxton G. Verification of dosimetric accuracy on the TrueBeam STx: rounded leaf effect of the high definition MLC. *Med Phys.* 2012 Oct;39(10):6360-71. Stanford University School of Medicine, Stanford, CA

Medical Advice Disclaimer

Varian as a medical device manufacturer cannot and does not recommend specific treatment or reimbursement approaches. Individual treatment results may vary.

Intended Use Summary

Varian Medical Systems' linear accelerators are intended to provide stereotactic radiosurgery and precision radiotherapy for lesions, tumors, and conditions anywhere in the body where radiation treatment is indicated.

Safety Statement

Varian Medical Systems' linear accelerators are intended to provide stereotactic radiosurgery and precision radiotherapy for lesions, tumors, and conditions anywhere in the body where radiation treatment is indicated.

Not all products or features are available for sale in all markets.

This bibliography is a comprehensive selection of articles but is not necessarily an exhaustive list of literature pertaining to Edge Radiosurgery.

varian

A Siemens Healthineers Company

varian.com

USA, Corporate Headquarters and Manufacturer

Varian Medical Systems, Inc.
Palo Alto, CA, USA
Phone +1-650-493 4000
800 544 4636
Fax +1-650-493 5637

EMEA and CIS Headquarters

Varian Medical Systems
International AG
Steinhausen, Switzerland
Phone +41-41-749 88 44
Fax +41-41-749 88 99

Asia Pacific Headquarters

Varian Medical Systems
Pacific, Inc.
Kowloon, Hong Kong,
SAR China
Phone +852-2724-2836
Fax +852-2369-4280

Australasian Headquarters

Varian Medical Systems
Australasia Pty Ltd.
Sydney, Australia
Phone +61-2-9485 0100
Fax +61-2-9485 0119

Latin American Headquarters

Varian Medical Systems
Brasil Ltda.
São Paulo, Brazil
Phone +55-11-3457 2655
Fax +55-11-3286 0034

© 2021 Varian Medical Systems, Inc. All rights reserved. Varian, Varian Medical Systems, and Edge are registered trademarks of Varian Medical Systems, Inc. All other trademarks are the property of their respective owners.