

## HD120 MLC™ High Definition Multileaf Collimator

### Bibliography

- Ohira S, Ueda Y, Kanayama N, Isono M, Inui S, Komiyama R, Washio H, Miyazaki M, Koizumi M, Teshima T, Konishi K. Impact of Multileaf Collimator Width on Dose Distribution in HyperArc Fractionated Stereotactic Irradiation for Multiple (-) Brain Metastases. *Anticancer Res*. 2021 Jun;41(6):3153-3159. Osaka International Cancer Institute, Osaka, Japan
- Bossart E, Mellon EA, Monterroso I, Elsayyad N, Diwanji T, Samuels S, Dogan N. Assessment of single isocenter linear accelerator radiosurgery for metastases and base of skull lesions. *Phys Med*. 2021 Jan;81:1-8. Sylvester Comprehensive Cancer Center, University of Miami, Miami, FL
- Taylor M, Williams J, Gleason JF Jr. Effects of Multileaf Collimator Design and Function When Using an Optimized Dynamic Conformal Arc Approach for Stereotactic Radiosurgery Treatment of Multiple Brain Metastases With a Single Isocenter: A Planning Study. *Cureus*. 2020 Aug 18;12(8):e9833. Alliance Cancer Care, Huntsville, AL
- Zhanerke Abisheva, Scott R. Floyd, Joseph K. Salama, John Kirkpatrick, Fang-Fang Yin, Michael J. Moravan, William Giles, Justus Adamson. The effect of MLC leaf width in single-isocenter multi-target radiosurgery with volumetric modulated arc therapy. *J Radiosurg SBRT*. 2019; 6(2): 131-138., Duke University Medical Center, Durham, NC
- Akino Y, Mizuno H, Tanaka Y, Isono M, Masai N, Yamamoto T. Inter-institutional variability of small-field-dosimetry beams among HD120™ multileaf collimators: a multi-institutional analysis. *Phys Med Biol*. 2018 Oct 18;63(20):205018. Osaka University Hospital, Suita, Osaka, Japan
- Woon WA, Ravindran PB, Ekayanake P, Lim YF. Validation of Delivery Consistency for Intensity-Modulated Radiation Therapy and Volumetric-Modulated Arc Therapy Plans. *J Med Phys*. 2018 Apr-Jun;43(2):119-128. The Brunei Cancer Center, Bandar Seri Begawan, Brunei Darussalam
- Hillman Y, Kim J, Chetty I, Wen N. Refinement of MLC modeling improves commercial QA dosimetry system for SRS and SBRT patient-specific QA. *Med Phys*. 2018 Apr;45(4):1351-1359. Henry Ford Health System, Detroit, MI
- Brito Delgado A, Cohen D, Eng TY, Stanley DN, Shi Z, Charlton M, Gutiérrez AN. Modeling the target dose fall-off in IMRT and VMAT planning techniques for cervical SBRT. *Med Dosim*. 2018 Spring;43(1):1-10. The University of Texas Health Science Center San Antonio, San Antonio, TX
- Mallory M, Pokhrel D, Badkul R, Jiang H, Lominska C, Wang F. Volumetric modulated arc therapy treatment planning of thoracic vertebral metastases using stereotactic body radiotherapy. *J Appl Clin Med Phys*. 2018 Mar;19(2):54-61. The University of Kansas Cancer Center, Kansas City, KS
- Han Z, Friesen S, Hacker F, Zygmanski P. Two-argument total scatter factor for small fields simultaneously collimated by MLC and jaws: application to stereotactic radiosurgery and radiotherapy. *Phys Med Biol*. 2017 Dec 14;63(1):015002. Brigham and Women's Hospital and Dana Farber Cancer Institute, Boston, MA
- Younge KC, Kuchta JR, Mikell JK, Rosen B, Bredfeldt JS, Matuszak MM. The impact of a high-definition multileaf collimator for spine SBRT. *J Appl Clin Med Phys*. 2017 Nov;18(6):97-103. University of Michigan Health System, Ann Arbor, MI
- Sasaki M, Tominaga M, Kamomae T, Ikushima H, Kitaoka M, Bando R, Sakuragawa K, Oita M. Influence of multi-leaf collimator leaf transmission on head and neck intensity-modulated radiation therapy and volumetric-modulated arc therapy planning. *Jpn J Radiol*. 2017 Sep;35(9):511-525. Tokushima University Hospital, Tokushima, Japan
- French SB, Bhagroo S, Nazareth DP, Podgorsak MB. Adapting VMAT plans optimized for an HD120 MLC for delivery with a Millennium MLC. *J Appl Clin Med Phys*. 2017 Sep;18(5):143-151. Roswell Park Cancer Institute, Buffalo, NY
- Gallagher KJ, Wong J, Zhang J. Photon caliper to achieve submillimeter positioning accuracy. *Phys Med Biol*. 2017 Sep 1;62(18):N404-N416. Oregon Health & Science University, Portland, OR
- Peng V, Suchowerska N, Rogers L, Claridge Mackonis E, Oakes S, McKenzie DR. Grid therapy using high definition multileaf collimators: realizing benefits of the bystander effect. *Acta Oncol*. 2017 Aug;56(8):1048-1059. University of Sydney, Camperdown, NSW, Australia

Kerolus MG, Sen N, Mayekar S, Templeton A, Turian J, Diaz A, Munoz L, Byrne RW, Sani S. Truebeam Radiosurgery for the Treatment of Trigeminal Neuralgia: Preliminary Results at a Single Institution. *Cureus.* 2017 Jun 16;9(6):e1362. Rush University Medical Center, Chicago, IL

Li J, Shi W, Andrews D, Werner-Wasik M, Lu B, Yu Y, Dicker A, Liu H. Comparison of Online 6 Degree-of-Freedom Image Registration of Varian TrueBeam Cone-Beam CT and BrainLab ExacTrac X-Ray for Intracranial Radiosurgery. *Technol Cancer Res Treat.* 2017 Jun;16(3):339-343. Thomas Jefferson University, Philadelphia, PA

Xu H, Brown S, Chetty IJ, Wen N. A Systematic Analysis of Errors in Target Localization and Treatment Delivery for Stereotactic Radiosurgery Using 2D/3D Image Registration. *Technol Cancer Res Treat.* 2017 Jun;16(3):321-331. Wayne State University, Detroit, MI

Gardner SJ, Lu S, Liu C, Wen N, Chetty IJ. Tuning of AcurosXB source size setting for small intracranial targets. *J Appl Clin Med Phys.* 2017 May;18(3):170-181. Henry Ford Health System, Detroit, MI

Oliver JA, Kelly P, Meeks SL, Willoughby TR, Shah AP. Orthogonal image pairs coupled with OSMS for noncoplanar beam angle, intracranial, single-isocenter, SRS treatments with multiple targets on the Varian Edge radiosurgery system. *Adv Radiat Oncol.* 2017 Apr 26;2(3):494-502. UF Health Cancer Center - Orlando Health, Orlando, FL

Chae SM, Lee KW, Son SH. Dosimetric impact of multileaf collimator leaf width according to sophisticated grade of technique in the IMRT and VMAT planning for pituitary adenoma lesion. *Oncotarget.* 2016 Nov 22;7(47):78119-78126. St. Mary's Hospital, Seoul, Korea

Chung JB, Kang SW, Eom KY, Song C, Choi KS, Suh TS. Comparison of Dosimetric Performance among Commercial Quality Assurance Systems for Verifying Pretreatment Plans of Stereotactic Body Radiotherapy Using Flattening-Filter-Free Beams. *J Korean Med Sci.* 2016 Nov;31(11):1742-1748. Seoul National University Bundang Hospital, Seongnam, Korea

Mancosu P, Fogliata A, Stravato A, Tomatis S, Cozzi L, Scorsetti M. Accuracy evaluation of the optical surface monitoring system on EDGE linear accelerator in a phantom study. *Med Dosim.* 2016 Summer;41(2):173-9. Humanitas Clinical and Research Center, Milan-Rozzano, Italy

Park JM, Park SY, Kim JH, Carlson J, Kim JI. The effect of extremely narrow MLC leaf width on the plan quality of VMAT for prostate cancer. *Radiat Oncol.* 2016 Jun 23;11:85. Seoul National University Hospital, Seoul, South Korea

Yang J, Tang G, Zhang P, Hunt M, Lim SB, LoSasso T, Mageras G. Dose calculation for hypofractionated volumetric-modulated arc therapy: approximating continuous arc delivery and tongue-and-groove modeling. *J Appl Clin Med Phys.* 2016 Mar 8;17(2):4989. Memorial Sloan-Kettering Cancer Center, New York, NY

Pokhrel D, McClinton C, Sood S, Badkul R, Saleh H, Jiang H, Lominska C. Monte Carlo evaluation of tissue heterogeneities corrections in the treatment of head and neck cancer patients using stereotactic radiotherapy. *J Appl Clin Med Phys.* 2016 Mar 8;17(2):6055. Univ of Kansas, Kansas City, KS

Tanny S, Sperling N, Parsai EI. Correction factor measurements for multiple detectors used in small field dosimetry on the Varian Edge radiosurgery system. *Med Phys.* 2015 Sep;42(9):5370-6. University of Toledo Medical Center, Toledo, OH

Wen N, Li H, Song K, Chin-Snyder K, Qin Y, Kim J, Bellon M, Gulam M, Gardner S, Doemer A, Devpura S, Gordon J, Chetty I, Siddiqui F, Ajlouni M, Pompa R, Hammoud Z, Simoff M, Kalkanis S, Movsas B, Siddiqui MS. Characteristics of a novel treatment system for linear accelerator-based stereotactic radiosurgery. *J Appl Clin Med Phys.* 2015 Jul 8;16(4):5313. Henry Ford Health System, Detroit, MI

Subramanian SV, Subramani V, Thirumalai Swamy S, Gandhi A, Chilukuri S, Kathirvel M. Is 5 mm MMLC suitable for VMAT-based lung SBRT? A dosimetric comparison with 2.5 mm HDMLC using RTOG-0813 treatment planning criteria for both conventional and high-dose flattening filter-free photon beams. *J Appl Clin Med Phys.* 2015 Jul 8;16(4):5415. Bharathiar University, India

Serna A, Puchades V, Mata F, Ramos D, Alcaraz M. Influence of multi-leaf collimator leaf width in radiosurgery via volumetric modulated arc therapy and 3D dynamic conformal arc therapy. *Phys Med.* 2015 May;31(3):293-6. Santa Lucia University Hospital, Murcia, Spain

Kothavade V, Jamema SV, Gupta T, Pungavkar S, Upasani M, Juvekar S, Jalali R. Which is the most optimal technique to spare hippocampus? - Dosimetric comparisons of SCRT, IMRT, and tomotherapy. *J Cancer Res Ther.* 2015 Apr-Jun;11(2):358-63. Global Hospital, Mumbai, India

Hernandez V, Abella R, Calvo JF, Jurado-Bruggemann D, Sancho I, Carrasco P. Determination of the optimal tolerance for MLC positioning in sliding window and VMAT techniques. *Med Phys.* 2015 Apr;42(4):1911-6. Hospital Sant Joan de Reus, Tarragona, Spain

Kim JI, Park SY, Kim HJ, Kim JH, Ye SJ, Park JM. The sensitivity of gamma-index method to the positioning errors of high-definition MLC in patient-specific VMAT QA for SBRT. *Radiat Oncol.* 2014 Jul 28;9:167. Seoul National University Hospital, Seoul, Korea

Bergman AM, Gete E, Duzenli C, Teke T. Monte Carlo modeling of HD120 multileaf collimator on Varian TrueBeam linear accelerator for verification of 6X and 6X FFF VMAT SABR treatment plans. *J Appl Clin Med Phys.* 2014 May 8;15(3):4686. BC Cancer Agency - Vancouver Centre, Vancouver, BC

Chae SM, Lee GW, Son SH. The effect of multileaf collimator leaf width on the radiosurgery planning for spine lesion treatment in terms of the modulated techniques and target complexity. *Radiat Oncol.* 2014 Mar 8;9(1):7. St. Mary's Hospital, Catholic University of Korea, Incheon, Korea

Hong CS, Ju SG, Kim M, Kim JI, Kim JM, Suh TS, Han Y, Ahn YC, Choi DH, Nam H, Park HC. Dosimetric effects of multileaf collimator leaf width on intensity-modulated radiotherapy for head and neck cancer. *Med Phys.* 2014 Feb;41(2):021712. Samsung Medical Center, Seoul, South Korea

Pommer T, Falk M, Poulsen PR, Keall PJ, O'Brien RT, Munck af Rosenschöld P. The impact of leaf width and plan complexity on DMLC tracking of prostate intensity modulated arc therapy. *Med Phys.* 2013 Nov;40(11):111717. Rigshospitalet, Copenhagen, Denmark

Shang Q, Qi P, Ferjani S, Xia P. Effect of MLC leaf width on treatment adaptation and accuracy for concurrent irradiation of prostate and pelvic lymph nodes. *Med Phys.* 2013 Jun;40(6):061701. Cleveland Clinic, Cleveland, OH

Tanyi JA, Doss EJ, Kato CM, Monaco DL, Zmeng L, Chen Y, Kubicky CD, Marquez CM, Fuss M. Dynamic conformal arc cranial stereotactic radiosurgery: implications of multileaf collimator margin on dose-volume metrics. *Br J Radiol.* 2012 Nov;85(1019):e1058-66. Oregon Health and Science University, Portland, OR

Kielar KN, Mok E, Hsu A, Wang L, Luxton G. Verification of dosimetric accuracy on the TrueBeam STx: rounded leaf effect of the high definition MLC. *Med Phys.* 2012 Oct;39(10):6360-71. Stanford University Medical Center, Palo Alto, CA

Vazquez-Quino LA, Massingill B, Shi C, Gutierrez A, Esquivel C, Eng T, Papanikolaou N, Stathakis S. Monte Carlo modeling of a Novalis Tx Varian 6 MV with HD-120 multileaf collimator. *J Appl Clin Med Phys.* 2012 Sep 6;13(5):3960. University of Texas Health Science Center at San Antonio, TX

Kim J, Wen N, Jin JY, Walls N, Kim S, Li H, Ren L, Huang Y, Doemer A, Faber K, Kunkel T, Balawi A, Garbarino K, Levin K, Patel S, Ajlouni M, Miller B, Nurushev T, Huntzinger C, Schulz R, Chetty IJ, Movsas B, Ryu S. Clinical commissioning and use of the Novalis Tx linear accelerator for SRS and SBRT. *J Appl Clin Med Phys.* 2012 May 10;13(3):3729. Henry Ford Health System, Detroit, MI

Wang L, Kielar KN, Mok E, Hsu A, Dieterich S, Xing L. An end-to-end examination of geometric accuracy of IGRT using a new digital accelerator equipped with onboard imaging system. *Phys Med Biol.* 2012 Feb 7;57(3):757-69. Stanford University Medical Center, Palo Alto, CA

Borges C, Zarza-Moreno M, Heath E, Teixeira N, Vaz P. Monte Carlo modeling and simulations of the High Definition (HD120) micro MLC and validation against measurements for a 6 MV beam. *Med Phys.* 2012 Jan;39(1):415. MedicalConsult SA, Lisboa, Portugal

Fujimoto S, Ono K, Furukawa K, Kudo T, Akagi Y, Koyama T, Hirokawa Y, Kawai S, Nakashima T. [Influence of high-definition multileaf collimator for three-dimensional conformal radiotherapy and intensity-modulated radiotherapy of prostate cancer]. *Nihon Hoshasen Gijutsu Gakkai Zasshi.* 2012;68(7):825-34. State of the Art Treatment Center, Hiroshima Heiwa Clinic, Hiroshima, Japan

Audet C, Poffenbarger BA, Chang P, Jackson PS, Lundahl RE, Ryu SI, Ray GR. Evaluation of volumetric modulated arc therapy for cranial radiosurgery using multiple noncoplanar arcs. *Med Phys.* 2011 Nov;38(11):5863-72. Palo Alto Medical Foundation, Palo Alto, CA

Ohtakara K, Hayashi S, Tanaka H, Hoshi H. Dosimetric comparison of 2.5 mm vs. 3.0 mm leaf width micro-multileaf collimator-based treatment systems for intracranial stereotactic radiosurgery using dynamic conformal arcs: implications for treatment planning. *Jpn J Radiol.* 2011 Nov;29(9):630-38. Gifu University Graduate School of Medicine, Gifu, Japan

Fix MK, Volken W, Frei D, Frauchiger D, Born EJ, Manser P. Monte Carlo implementation, validation, and characterization of a 120 leaf MLC. *Med Phys.* 2011 Oct;38(10):5311-20. Inselspital and University of Bern, Bern, Switzerland

Tanyi JA, Kato CM, Chen Y, Chen Z, Fuss M. Impact of the high-definition multileaf collimator on linear accelerator-based intracranial stereotactic radiosurgery. *Br J Radiol.* 2011 Jul;84(1003):629-38. Oregon Health and Science University, Portland, OR

Sharma DS, Dongre PM, Mhatre V, Heigrujam M. Physical and dosimetric characteristic of high-definition multileaf collimator (HDMLC) for SRS and IMRT. *J Appl Clin Med Phys.* 2011 Apr 14;12(3):3475. Kokilaben Dhirubhai Ambani Hospital, Mumbai, India

Dhabaan A, Elder E, Schreiber E, Crocker I, Curran WJ, Oyesiku NM, Shu HK, Fox T. Dosimetric performance of the new high-definition multileaf collimator for intracranial stereotactic radiosurgery. *J Appl Clin Med Phys.* 2010 Jun 21;11(3):3040. Emory University, Atlanta, GA

Clift C, Thomas A, Adamovics J, Chang Z, Das I, Oldham M. Toward acquiring comprehensive radiosurgery field commissioning data using the PRESAGE/optical-CT 3D dosimetry system. *Phys Med Biol.* 2010 Mar 7;55(5):1279-93. Duke University Medical Center, Durham, NC

Tanyi JA, Summers PA, McCracken CL, Chen Y, Ku LC, Fuss M. Implications of a high-definition multileaf collimator (HD-MLC) on treatment planning techniques for stereotactic body radiation therapy (SBRT): a planning study. *Radiat Oncol.* 2009 Jul 10;4:22. Oregon Health & Science University, Portland, OR

Wu QJ, Wang Z, Kirkpatrick JP, Chang Z, Meyer JJ, Lu M, Huntzinger C, Yin FF. Impact of collimator leaf width and treatment technique on stereotactic radiosurgery and radiotherapy plans for intra- and extracranial lesions. *Radiat Oncol.* 2009 Jan 21;4(1):3. Duke University, Durham, NC

Wuu CS, Kessel J, Xu Y. 3-D doseimetric evaluation of 2.5 mm HD120 multileaf system for intensity modulated stereotactic radiosurgery using optical CT based gel dosimetry. *J. Phys.: Conf. Ser.* 2009 164(1):2-6. Columbia University, New York, NY

Fogliata A, Clivio A, Nicolini G, Vanetti E, Cozzi L. Intensity modulation with photons for benign intracranial tumours: a planning comparison of volumetric single arc, helical arc and fixed gantry techniques. *Radiother Oncol.* 2008 Dec;89(3):254-62. Oncology Institute of Southern Switzerland, Bellinzona, Switzerland

Chang Z, Wang Z, Wu QJ, Yan H, Bowsher J, Zhang J, Yin FF. Dosimetric characteristics of novalis Tx system with high definition multileaf collimator. *Med Phys.* 2008 Oct;35(10):4460-3. Duke University Medical Center, Durham, NC

#### Medical Advice Disclaimer

Varian as a medical device manufacturer cannot and does not recommend specific treatment or reimbursement approaches. Individual treatment results may vary.

#### Intended Use Summary

Varian Medical Systems' linear accelerators are intended to provide stereotactic radiosurgery and precision radiotherapy for lesions, tumors, and conditions anywhere in the body where radiation treatment is indicated.

#### Safety Statement

Varian Medical Systems' linear accelerators are intended to provide stereotactic radiosurgery and precision radiotherapy for lesions, tumors, and conditions anywhere in the body where radiation treatment is indicated.

*Not all products or features are available for sale in all markets.*

*This bibliography is a comprehensive selection of articles but is not necessarily an exhaustive list of literature pertaining to HD120 MLC™.*

**varian**

A Siemens Healthineers Company

varian.com

#### USA, Corporate Headquarters and Manufacturer

Varian Medical Systems, Inc.  
Palo Alto, CA, USA  
Phone +1-650-493 4000  
800 544 4636  
Fax +1-650-493 5637

#### EMEA and CIS Headquarters

Varian Medical Systems  
International AG  
Steinhausen, Switzerland  
Phone +41-41-749 88 44  
Fax +41-41-749 88 99

#### Asia Pacific Headquarters

Varian Medical Systems  
Pacific, Inc.  
Kowloon, Hong Kong,  
SAR China  
Phone +852-2724-2836  
Fax +852-2369-4280

#### Australasian Headquarters

Varian Medical Systems  
Australasia Pty Ltd.  
Sydney, Australia  
Phone +61-2-9485 0100  
Fax +61-2-9485 0119

#### Latin American Headquarters

Varian Medical Systems  
Brasil Ltda.  
São Paulo, Brazil  
Phone +55-11-3457 2655  
Fax +55-11-3286 0034

© 2021 Varian Medical Systems, Inc. All rights reserved. Varian, Varian Medical Systems and HD120 MLC are registered trademarks of Varian Medical Systems, Inc. All other trademarks are the property of their respective owners.