

Halcyon® System Bibliography*

CLINICAL SECTION

Brain

Reynoso FJ, Hugo GD, Mutic S, Gach HM, Knutson NC. Lateral head flexion as a noncoplanar solution for ring gantry stereotactic radiosurgery. *Med Phys.* 2020 Mar;47(3):1181-1188. Washington University School of Medicine in St Louis, St Louis, MO

Flores-Martinez E, Cerviño LI, Pawlicki T, Kim GY. Assessment of the use of different imaging and delivery techniques for cranial treatments on the Halcyon linac. *J Appl Clin Med Phys.* 2020 Jan;21(1):53-61. University of California, San Diego, La Jolla, CA

Li T, Irmén P, Liu H, Shi W, Alonso-Basanta M, Zou W, Teo BK, Metz JM, Dong L. Dosimetric Performance and Planning/Delivery Efficiency of a Dual-Layer Stacked and Staggered MLC on Treating Multiple Small Targets: A Planning Study Based on Single-Isocenter Multi-Target Stereotactic Radiosurgery (SRS) to Brain Metastases. *Front Oncol.* 2019 Jan 22;9:7. University of Pennsylvania, Philadelphia, PA

Spine and Spine SBRT

Petroccia HM, Malajovich I, Barsky AR, Ghiam AF, Jones J, Wang C, Zou W, Teo BK, Dong L, Metz JM, Li T. Spine SBRT With Halcyon™: Plan Quality, Modulation Complexity, Delivery Accuracy, and Speed. *Front Oncol.* 2019 Apr 26;9:319. University of Pennsylvania, Perelman Center for Advanced Medicine, Philadelphia, PA

Pokhrel D, Stephen J, Webster A, Bernard ME. Double-vertebral segment SBRT via novel ring-mounted Halcyon Linac: Plan quality, delivery efficiency and accuracy. *Med Dosim.* 2022 Spring;47(1):20-25. University of Kentucky, Lexington, KY

Li F, Park J, Lalonde R, Jang SY, diMayorca MS, Flickinger JC, Keller A, Huq MS. Is Halcyon feasible for single thoracic or lumbar vertebral segment SBRT? *J Appl Clin Med Phys.* 2022 Jan;23(1):e13458. UPMC Hillman Cancer Center, Pittsburgh, Pennsylvania, USA

Head and Neck

Riley C, Cox C, Graham S, Havran H, Kramer B, Netherton T, Peterson C, Williamson T, Court L. Varian Halcyon™ dosimetric comparison for multiarc VMAT prostate and head-and-neck cancers. *Med Dosim.* 2018 Sep 6. pii: S0958-3947(18)30084-0. The University of Texas MD Anderson Cancer Center, Houston, TX and University of Pennsylvania, Philadelphia, PA [Epub ahead of print]

Li T, Scheuermann R, Lin A, Teo BK, Zou W, Swisher-McClure S, Alonso-Basanta M, Lukens JN, Fotouhi Ghiam A, Kennedy C, Kim MM, Mihailidis D, Metz JM, Dong L. Impact of Multi-leaf Collimator Parameters on Head and Neck Plan Quality and Delivery: A Comparison between Halcyon™ and TrueBeam® Treatment Delivery Systems. *Cureus.* 2018 Nov 28;10(11):e3648. University of Pennsylvania, Philadelphia, PA

Michiels S, Poels K, Crijns W, Delombaerde L, De Roover R, Vanstraelen B, Haustermans K, Nuyts S, Depuydt T. Volumetric modulated arc therapy of head-and-neck cancer on a fast-rotating O-ring linac: Plan quality and delivery time comparison with a C-arm linac. *Radiother Oncol.* 2018 Sep;128(3):479-484. University of Leuven, Belgium

Breast

Morris R, Laugeman E, Hilliard J, Zoberi I, Heerman A, Hugo G, Mutic S, Cai B. Field-in-field breast planning for a jawless, double-stack MLC LINAC using flattening-filter-free beams. *J Appl Clin Med Phys.* 2019 Nov;20(11):14-26. Washington University, St. Louis, MO

O'Grady F, Barsky AR, Anamalayil S, Freedman GM, Kennedy C, Cai B, Laugeman E, Dong L, Hugo GD, Metz JM, Mutic S, Taunk NK, Li T. Increase in Superficial Dose in Whole-Breast Irradiation With Halcyon Straight-Through Linac Compared With Traditional C-arm Linac With Flattening Filter: In vivo Dosimetry and Planning Study. *Adv Radiat Oncol.* 2019 Aug 21;5(1):120-126. University of Pennsylvania, Perelman Center for Advanced Medicine, Philadelphia, PA & Washington University School of Medicine in St Louis, St Louis, MO

Flores-Martinez E, Kim GY, Yashar CM, Cerviño LI. [Dosimetric study of the plan quality and dose to organs at risk on tangential breast treatments using the Halcyon™ linac.](#) *J Appl Clin Med Phys.* 2019 Jul;20(7):58-67. University of California, San Diego, La Jolla, CA

Barsky AR, O'Grady F, Kennedy C, Taunk NK, Dong L, Metz JM, Li T, Freedman GM. [Initial Clinical Experience Treating Patients with Breast Cancer on a 6-MV Flattening-Filter-Free O-Ring Linear Accelerator.](#) *Adv Radiat Oncol.* 2019 Jun 8;4(4):571-578. University of Pennsylvania, Perelman Center for Advanced Medicine, Philadelphia, PA

Kim MM, Kennedy C, Scheuermann R, Freedman G, Li T. [Whole Breast and Lymph Node Irradiation using Halcyon™ 2.0 Utilizing Automatic Multi-isocenter Treatment Delivery and Daily Kilovoltage Cone-beam Computed Tomography.](#) *Cureus.* 2019 May 24;11(5):e4744. University of Pennsylvania, Philadelphia, PA

Kennedy C, Freedman G, Taunk N, Scheuermann R, Dong L, Metz JM, Li T. [Whole Breast Irradiation with Halcyon™ 2.0: Workflow and Efficiency of Field-in-Field Treatment with Dynamic Beam Flattening Technique and kV Cone Beam Computed Tomography.](#) *Cureus.* 2018 Oct 29;10(10):e3510. University of Pennsylvania, Philadelphia, PA

Richter A, Wegener S, Breuer K, Razinskas G, Weick S, Exner F, Bratengeier K, Flentje M, Sauer O, Polat B. [Comparison of sliding window and field-in-field techniques for tangential whole breast irradiation using the Halcyon and Synergy Agility systems.](#) *Radiat Oncol.* 2021 Nov 6;16(1):213. University of Wuerzburg, Wuerzburg, Germany

Pelvis

Panda S, Swamidas J, Chopra S, Mangaj A, Fogliata A, Kupelian P, Agarwal JP, Cozzi L. [Treatment planning comparison of volumetric modulated arc therapy employing a dual-layer stacked multi-leaf collimator and helical tomotherapy for cervix uteri.](#) *Radiat Oncol.* 2020 Jan 30;15(1):22. Tata Memorial Centre, Navi Mumbai, Maharashtra, India

Jarema T, Aland T. [Using the iterative kV CBCT reconstruction on the Varian Halcyon linear accelerator for radiation therapy planning for pelvis patients.](#) *Phys Med.* 2019 Dec;68:112-116. Icon Cancer Centres, Brisbane, AU

Kim H, Huq MS, Lalonde R, Houser CJ, Beriwal S, Heron DE. [Early clinical experience with varian halcyon V2 linear accelerator: Dual-isocenter IMRT planning and delivery with portal dosimetry for gynecological cancer treatments.](#) *J Appl Clin Med Phys.* 2019 Nov;20(11):111-120 University of Pittsburgh School of Medicine, Pittsburgh, PA

Li C, Chen J, Zhu J, Gong G, Tao C, Li Z, Lu J, Yin Y. [Plan quality comparison for cervical carcinoma treated with Halcyon and Trilogy intensity-modulated radiotherapy.](#) *J Cancer.* 2019 Oct 15;10(24):6135-6141. Shandong Cancer Hospital and Institute, Jinan, China

Prostate and Prostate SBRT

Riley C, Cox C, Graham S, Havran H, Kramer B, Netherton T, Peterson C, Williamson T, Court L. [Varian Halcyon™ dosimetric comparison for multiarc VMAT prostate and head-and-neck cancers.](#) *Med Dosim.* 2018 Sep 6. pii: S0958-3947(18)30084-0. The University of Texas MD Anderson Cancer Center, Houston, TX and University of Pennsylvania, Philadelphia, PA [Epub ahead of print]

Pokhrel D, Tackett T, Stephen J, Visak J, Amin-Zimmerman F, McGregor A, Strup SE, St Clair W. [Prostate SBRT using O-Ring Halcyon Linac - Plan quality, delivery efficiency, and accuracy.](#) *J Appl Clin Med Phys.* 2021 Jan;22(1):68-75. University of Kentucky, Lexington, KY

De Roover R, Crijns W, Poels K, Dewit B, Draulans C, Haustermans K, Depuydt T. [Automated treatment planning of prostate stereotactic body radiotherapy with focal boosting on a fast-rotating O-ring linac: Plan quality comparison with C-arm linacs.](#) *J Appl Clin Med Phys.* 2021 Sep;22(9):59-72. University Hospitals Leuven, Leuven, Belgium

Lung

Pokhrel D, Visak J, Critchfield LC, Stephen J, Bernard ME, Randall M, Kudrimoti M. [Clinical validation of ring-mounted halcyon linac for lung SBRT: comparison to SBRT-dedicated C-arm linac treatments.](#) *J Appl Clin Med Phys.* 2021 Jan;22(1):261-270. University of Kentucky, Lexington, KY

Total Body/Long Fields

Uehara T, Monzen H, Tamura M, Inada M, Otsuka M, Doi H, Matsumoto K, Nishimura Y. [Feasibility study of volumetric modulated arc therapy with Halcyon™ linac for total body irradiation.](#) *Radiat Oncol.* 2021 Dec 14;16(1):236. Kindai University, Osaka, Japan

GENERAL SECTION

Global Health and Economics

Velarde A, Najera KD, Gay H, Powderly WG, Mutic S, Green J, Michalski JM, Henke L, de Falla V, Laugeman E, Catu M, Hugo GD, Cai B, van Rheeën J. Taking Guatemala From Cobalt to IMRT: A Tale of US Agency Collaboration With Academic Institutions and Industry. *Int J Radiat Oncol Biol Phys.* 2020 Aug 1;107(5):867-872. Liga Nacional Contra El Cáncer e Instituto de Cancerología-INCAN, Guatemala City, Guatemala, Washington University, St Louis, Missouri, and National Institutes of Health, Bethesda, Maryland

TECHNICAL SECTION

Planning and Quality Assurance

Tamura M, Matsumoto K, Otsuka M, Monzen H. Plan complexity quantification of dual-layer multi-leaf collimator for volumetric modulated arc therapy with Halcyon linac. *Phys Eng Sci Med.* 2020 Jul 9. doi: 10.1007/s13246-020-00891-2. Kindai University, Osaka, Japan. [Epub ahead of print]

Fogliata A, Cayez R, Garcia R, Khamphan C, Reggiori G, Scorsetti M, Cozzi L. Technical Note: Flattening filter free beam from Halcyon linac: Evaluation of the profile parameters for quality assurance. *Med Phys.* 2020 May 4. doi: 10.1002/mp.14217. Humanitas Clinical and Research Center, Milan, Rozzano, Italy. [Epub ahead of print]

Matsumoto K, Otsuka M, Tamura M, Monzen H, Okumura M. Comparison of O-ring and General Linacs for Treatment Planning of Volumetric Modulated Arc Therapy. *Nihon Hoshasen Gijutsu Gakkai Zasshi.* 2020;76(4):339-345. Kindai University Hospital, Osaka, Japan

Laugeman E, Heermann A, Hilliard J, Watts M, Roberson M, Morris R, Goddu S, Sethi A, Zoberi I, Kim H, Mutic S, Hugo G, Cai B. Comprehensive validation of halcyon 2.0 plans and the implementation of patient specific QA with multiple detector platforms. *J Appl Clin Med Phys.* Jul;21(7):39-48. Washington University School of Medicine in St Louis, St Louis, MO

Li Y, Netherton T, Nitsch PL, Gao S, Klopp AH, Balter PA, Court LE. Independent validation of machine performance check for the Halcyon and TrueBeam linacs for daily quality assurance. *J Appl Clin Med Phys.* 2018 Sep;19(5):375-382. The University of Texas MD Anderson Cancer Center, Houston, TX

Gao S, Chetvertkov MA, Cai B, Dwivedi A, Mihailidis D, Ray X, Netherton T, Court LE, Simon WE, Balter PA. Beam energy metrics for the acceptance and quality assurance of Halcyon linear accelerator. *J Appl Clin Med Phys.* 2021 Jul;22(7):121-127. The University of Texas MD Anderson Cancer Center, Houston, TX

Cozzi L, Fogliata A, Thompson S, Franzese C, Franceschini D, de Rose F, Tomatis S, Scorsetti M. Critical Appraisal of the Treatment Planning Performance of Volumetric Modulated Arc Therapy by Means of a Dual Layer Stacked Multileaf Collimator for Head and Neck, Breast, and Prostate. *Technol Cancer Res Treat.* 2018 Jan 1;17:1533033818803882. Humanitas Research Hospital, Milan-Rozzano, Italy

Accuracy

Huang Y, Du Y, Li C, Wang H, Zu Z, Feng Z, Zhou S, Wu H, Zhang Y. Pediatric cone beam CT on Varian Halcyon™ and TrueBeam® radiotherapy systems: radiation dose and positioning accuracy evaluations. *J Radiol Prot.* 2019 Sep;39(3):739-748 Peking University Cancer Hospital & Institute, Beijing, People's Republic of China

Petroccia HM, Malajovich I, Barsky AR, Ghiam AF, Jones J, Wang C, Zou W, Teo BK, Dong L, Metz JM, Li T. Spine SBRT With Halcyon™: Plan Quality, Modulation Complexity, Delivery Accuracy, and Speed. *Front Oncol.* 2019 Apr 26;9:319. University of Pennsylvania, Perelman Center for Advanced Medicine, Philadelphia, PA

De Roover R, Crijns W, Poels K, Michiels S, Nulens A, Vanstraelen B, Petillion S, De Brabandere M, Haustermans K, Depuydt T. Validation and IMRT/VMAT delivery quality of a preconfigured fast-rotating O-ring linac system. *Med Phys.* 2019 Jan;46(1):328-339. University Hospitals Leuven, Leuven, Belgium

Wang H, Huang Y, Hu Q, Li C, Liu H, Wang X, Li W, Ma W, Pu Y, Du Y, Wu H, Zhang Y. A Simulated Dosimetric Study of Contribution to Radiotherapy Accuracy by Fractional Image Guidance Protocol of Halcyon System. *Front Oncol.* 2021 Jan 25;10:543147. Peking University Cancer Hospital & Institute, Beijing, China

Imaging

Huang Y, Li C, Wang H, Hu Q, Wang R, Chang C, Ma W, Li W, Wu H, Zhang Y. A quantitative evaluation of deformable image registration based on MV cone beam CT images: Impact of deformation magnitudes and image modalities. *Phys Med.* 2020 Mar;71:82-87. Peking University Cancer Hospital & Institute, Beijing, China

Wang R, Du Y, Yao K, Liu Z, Wang H, Yue H, Zhang Y, Wu H. Halcyon clinical performance evaluation: A log file-based study in comparison with a C-arm Linac. *Phys Med*. 2020 Mar;71:14-23. Peking University Cancer Hospital & Institute, Beijing, China.

Ray X, Bojecho C, Moore KL. Evaluating the sensitivity of Halcyon's automatic transit image acquisition for treatment error detection: A phantom study using static IMRT. *J Appl Clin Med Phys*. 2019 Nov;20(11):131-143. University of California, San Diego, La Jolla, CA

Gao S, Netherton T, Chetvertkov MA, Li Y, Court LE, Simon WE, Shi J, Balter PA. Acceptance and verification of the Halcyon-Eclipse linear accelerator-treatment planning system without 3D water scanning system. *J Appl Clin Med Phys*. 2019 Oct;20(10):111-117. The University of Texas MD Anderson Cancer Center, Houston, TX

Netherton T, Li Y, Nitsch P, Shaitelman S, Balter P, Gao S, Klopp A, Muruganandham M, Court L. Interplay effect on a 6-MV flattening-filter-free linear accelerator with high dose rate and fast multi-leaf collimator motion treating breast and lung phantoms. *Med Phys*. 2018 Jun;45(6):2369-2376 University of Texas Graduate School of Biomedical Sciences and The University of Texas MD Anderson Cancer Center, Houston, TX

Dosimetry & Characterization

Bollinger D, Laugeman E, Li T, Hilliard J, Heermann A, Kim H, Hugo G, Mutic S, Dong L, Cai B. Technical Note: Dosimetric characterization of the dynamic beam flattening MLC sequence on a ring shaped, Jawless Linear Accelerator with double stacked MLC. *Med Phys*. 2020 Mar;47(3):948-957. University of Pennsylvania, Perelman Center for Advanced Medicine, Philadelphia, PA & Washington University School of Medicine in St Louis, St Louis, MO

Gay SS, Netherton TJ, Cardenas CE, Ger RB, Balter PA, Dong L, Mihailidis D, Court LE. Dosimetric impact and detectability of multi-leaf collimator positioning errors on Varian Halcyon. *Med Phys*. 2019 Aug;20(8):47-55. The University of Texas MD Anderson Cancer Center, Houston, TX

Malajovich I, Teo BK, Petrocchia H, Metz JM, Dong L, Li T. Characterization of the Megavoltage Cone-Beam Computed Tomography (MV-CBCT) System on Halcyon™ for IGRT: Image Quality Benchmark, Clinical Performance, and Organ Doses. *Front Oncol*. 2019 Jun 12;9:496. University of Pennsylvania, Philadelphia, PA

Lim TY, Dragojević I, Hoffman D, Flores-Martinez E, Kim GY. Characterization of the Halcyon™ multileaf collimator system. *J Appl Clin Med Phys*. 2019 Apr;20(4):106-114. University of California, San Diego, La Jolla, CA

Cai B, Laugeman E, Mazur TR, Park JC, Henke LE, Kim H, Hugo GD, Mutic S, Li H. Characterization of a prototype rapid kilovoltage x-ray image guidance system designed for a ring shape radiation therapy unit. *Med Phys*. 2019 Mar;46(3):1355-1370. Washington University, St. Louis, MO

Huang Y, Wu H, Hu Q, Wang M, Wang R, Zu Z, Wang H, Wang W, Zhang Y. Dose calculation deviations induced by fractional image-guided-couch-shifts for Varian Halcyon MV cone beam CT. *Phys Med*. 2019 Feb;58:66-71. Peking University Cancer Hospital & Institute, Beijing, China

Lloyd SAM, Lim TY, Fave X, Flores-Martinez E, Atwood TF, Moiseenko V. TG-51 reference dosimetry for the Halcyon™: A clinical experience. *J Appl Clin Med Phys*. 2018 Jul;19(4):98-102. University of California, San Diego, La Jolla, CA

Li Y, Netherton T, Nitsch PL, Balter PA, Gao S, Klopp AH, Court LE. Normal tissue doses from MV image-guided radiation therapy (IGRT) using orthogonal MV and MV-CBCT. *J Appl Clin Med Phys*. 2018 May;19(3):52-57. The University of Texas MD Anderson Cancer Center, Houston, TX

Aland T, Jarema T, Spalding M, Kairn T, Trapp J. Use of in vivo transit portal images to detect gross inter-fraction patient geometry changes on an O-ring type linear accelerator for pelvis and head/neck patients. *Biomed Phys Eng Express*. 2021 Oct 25;7(6). Icon Cancer Centres, Brisbane, Australia.

Bojecho C, Nelson T, Simpson D, Moiseenko V. Assessment of predictive indicators of acute gastrointestinal toxicity using in vivo transmission images. *Med Phys*. 2021 Dec;48(12):8152-8162. University of California San Diego, La Jolla, CA

Van Esch A, Kulmala A, Rochford R, Kauppinen J, Harju A. Testing of an enhanced leaf model for improved dose calculation in a commercial treatment planning system. *Med Phys*. 2022 Oct 3. 7Sigma, QA-team in radiotherapy physics, Tildonk, Belgium

Commissioning

Netherton T, Li Y, Gao S, Klopp A, Balter P, Court LE, Scheuermann R, Kennedy C, Dong L, Metz J, Mihailidis D, Ling C, Young Lee M, Constantin M, Thompson S, Kauppinen J, Uusitalo P. Experience in commissioning the Halcyon™ linac. *Med Phys.* 2019 Oct;46(10):4304-4313. The University of Texas MD Anderson Cancer Center, Houston, TX

Pawlicki T, Atwood T, McConnell K, Kim GY. Clinical safety assessment of the Halcyon™ system. *Med Phys.* 2019 Oct;46(10):4340-4345. University of California, San Diego, La Jolla, CA

Teo PT, Hwang MS, Shields WG, Kosterin P, Jang SY, Heron DE, Lalonde RJ, Huq MS. Application of TG-100 risk analysis methods to the acceptance testing and commissioning process of a Halcyon™ linear accelerator. *Med Phys.* 2019 Mar;46(3):1341-1354. University of Pittsburgh School of Medicine and UPMC Hillman Cancer Center, Pittsburgh, PA

Cai B, Laugeman E, Hsu H, Green O, Knutson N, Goddu SM, Mutic S, Du S, Henke L, Kim H, Hugo GD. Technical Note: Self-shielding evaluation and radiation leakage measurement of a jawless ring gantry linac with a beam stopper. *Med Phys.* 2021 Jun;48(6):3143-3150. Washington University, St. Louis, MO

**This bibliography is a comprehensive selection of articles but is not necessarily an exhaustive list of literature pertaining to the Halcyon® system.*

Intended Use Summary

Varian Medical Systems' linear accelerators are intended to provide stereotactic radiosurgery and precision radiotherapy for lesions, tumors, and conditions anywhere in the body where radiation treatment is indicated.

Safety Statement

Radiation treatments may cause side effects that can vary depending on the part of the body being treated. The most frequent ones are typically temporary and may include, but are not limited to, irritation to the respiratory, digestive, urinary or reproductive systems, fatigue, nausea, skin irritation, and hair loss. In some patients, they can be severe. Treatment sessions may vary in complexity and time. Radiation treatment is not appropriate for all cancers.

varian

A Siemens Healthineers Company

varian.com

USA, Corporate Headquarters and Manufacturer

Varian Medical Systems, Inc.
3100 Hansen Way
Palo Alto, CA 94304
Tel: 650.424.5700
800.544.4636

Headquarters Europe, Eastern Europe, Middle & Near East, India, Africa

Siemens Healthineers International AG
Steinhausen, Switzerland
Tel: 41.41.749.8844

Asia Pacific Headquarters

Varian Medical Systems Pacific, Inc.
Kowloon, Hong Kong
Tel: 852.2724.2836

Australasian Headquarters

Varian Medical Systems Australasia Pty Ltd.
Sydney, Australia
Tel: 61.2.9485.0100

Latin American Headquarters

Varian Medical Systems Brasil Ltda.
São Paulo, Brazil
Tel: 55.11.3457.2655